

R E M A R K S

By this Amendment the specification has been amended to further describe the inventive X-ray-sensitive camera (page 9) and to correct an insert previously proposed (page 13), claims 28 and 29 have been canceled, claim 31 has been amended to further define the inventive X-ray-sensitive camera, claims 33 and 34 have been revised to address formality issues, and new claim 36 has been added to define a further feature. Entry is requested.

Note the concurrently filed Submission of Replacement Drawing Sheets.


In the outstanding Office Action the examiner has rejected claim 31 under 35 U.S.C. 102(b) as being anticipated by Zeller et al.

This rejection cannot apply to amended claim 31. Zeller et al. disclose an X-ray detector camera which includes first and second detectors 18, 18'; however, there is no disclosure of positioning one detector behind the other with X-ray-sensitive surfaces facing in opposite directions (in Figs. 8 and 9 the X-ray-sensitive surfaces face in the same direction).

Favorable reevaluation is requested.

Respectfully submitted,

By:

A handwritten signature in black ink, appearing to read "Richard H. Tushin", written over a horizontal line.

Richard H. Tushin
Registration No. 27,297
Franklin Square, Third Floor West
1300 I Street, N.W.
Washington, DC 20005-3353
(202) 906-8680